

Certificate of Assurance

This is to certify that the product has undergone following tests and analysis and meets the Axiva Quality Testing System approved by an accredited registered body to an ISO 9001:2015 Quality System Standards and supplied in accordance with a Quality Management System governed by Axiva Sichem Biotech

Product	: Conjugate Release Pad	Cat. No.	: 16613T01
Lot No.	: GP67373100	Size (mm)	: 200 x 300

Parameter	Criteria	status
Material of Construction	Polyester	Passed
Thickness	350 - 440 μ m	Passed
Weight	9 - 11 mg/cm ²	Passed
Wicking Time	40 sec. \pm 15 sec Time taken to travel 4 cm distance by distilled water	Passed
Colour	White	Passed
Visual Inspection	Pad surface uniform and free from particles and coloured spots.	Passed

Shelf life : For best performance, please use before 19 Oct. 2022 , if Kept in original packing.

Storage : Please store the Pad in original packing at 15°C to 30°C. Material should be kept away from heat radiators and moisture.



Head of Quality Control

Issue Date: 19 Oct. 2020

Axiva Sichem Biotech