

Certificate of Assurance

This is to certify that the product has undergone following tests and analysis and meets the Axiva Quality Testing System approved by an accredited registered body to an ISO 9001:2015 Quality System Standards and supplied in accordance with a Quality Management System governed by Axiva Sichem Biotech

Product	: Flow Through Membrane	Pore Size (μm)	: 0.45
Lot No.	: FT00808070	Size (mm)	: 100 x 100

Parameter	Criteria	Status
Material	Cellulose Nitrate	Passed
Thickness	450 – 550 μm	Passed
Bubble Point	33 – 40 Psi	Passed
Water Flow Rate	≥ 20 ml/min@10psi	Passed
Wetting Test	Wetting Time ≤ 5 Sec and Round Shape (With 10 μl of 0.3% BSA in PBS 0.01M pH 7.4)	Passed
Colour	White	Passed
Visual Inspection	Free from particles, coloured spots and visual surface deformity	Passed

Shelf life : For best performance, please use before 28 Jul. 2022, if kept in original packing.

Storage : Please store the membrane in original packing at 15°C to 24°C. Please avoid direct sunlight and should be kept away from heat radiators and dust.



Head of Quality Control

Issue Date: 28 Jul. 2020